



HPSEDC/PCB/CAAQMS/2K19-

Date: 04/06/2019

CORRIGENDUM-II


Tender No. : HPSEDC/PCB/CAAQMS/2K19-


HPSEDC published a Tender Notice from globally/ internationally eligible bidders for Supply, Installation, Commissioning And Operation & Maintenance Services For Continuous Ambient Air Quality Monitoring Stations (CAAQMS) at Shimla and related services as specified in the tender document after getting approval from Election Department.


As discussed in the pre-bid meeting, following Sections/Clauses/ Sub Clause may be read as under:


| Sr. No. | Section/ Clause /Sub clause/ page No | As per tender document | May be read as |
|---------|--------------------------------------|--|--|
| 1. | Clause-2.5.4 (ii), Vol.-1, Page-12 | The bidder, as authorized representative, supplied, tested and commissioned minimum six (6) nos. similar air monitoring stations either container based or otherwise in any one year of the last three years which should be in satisfactory operation with no adverse report for last two (2) year as on date of bid opening. | Complete package is required as mentioned in the tender document. |
| 2. | Note-2, Vol.-1, Page-94 | Quoted cost of O&M of a CAAQMS for each year should be at least 12% of the cost of supply & installation of the CRTAQMS. in case the bidder quotes O&M cost lower than 15% of the supply & installation cost, the owner will reduce the supply & installation cost suitably, so as to make the O&M cost as 15% of the supply & installation cost, keeping the total quoted cost as unchanged | Quoted cost of O&M of a CAAQMS for each year should be at least 15% of the cost of supply & installation of the CRTAQMS. in case the bidder quotes O&M cost lower than 15% of the supply & installation cost, the owner will reduce the supply & installation cost suitably, so as to make the O&M cost as 15% of the supply & installation cost, keeping the total quoted cost as unchanged |
| 3. | BOQ, Vol.-1, Page-95, 96 | Attachment-2 SUMMARY OF BID PRICE Attachment-3, BID PRICE BREAKUP FOR EQUIPMENT | Based on the SBI currency rates at the time of bid submission, bidder can submit price bid/ BOQ prices after converting multiple currency in Indian Rupee. Any discount & |

| | | | |
|----|---|---|---|
| | | (CIF) PRICE Attachment-3A, BID PRICE BREAKUP FOR O&M OF CAAQMS FOR FIVE YEARS | rebate in taxes may also be conveyed by the bidder in technical bid. Bid currency conversion rates should also be mentioned in the technical bid. |
| 4. | Technical Specification, Vol.-II, Page-11 | 2.1 SPLIT AIR CONDITIONER (2.0 TON CAPACITY) 2.1.1 Type & Capacity: 2 Nos. split type, 2 ton capacity AC, roof mounted of 5 star rating with an automatic timer. Separate Automatic Voltage stabilizer will be provided with each unit. | 2.1 SPLIT AIR CONDITIONER (2.0 TON CAPACITY) 2.1.1 Type & Capacity: 2 Nos. split type, 2 ton capacity AC, roof mounted Inverter compressor split AC with an automatic timer. Separate Automatic Voltage stabilizer will be provided with each unit. |


Kuldeep Rana,
Data Controller,
HPSEDC


Suraj Parkash,
Data Controller,
HPSEDC


Arun Dev Sharm,
Dy. Manager, (P&A)
HPSEDC


Rajeev Shirma,
General Manager,
(Technical), HPSEDC